

- e-CHECK*** (XS) Quality Control for the **ProCyte Dx*** Analyzer
- e-CHECK*** (XS) contrôle qualité pour l'analyseur **ProCyte Dx***
- e-CHECK*** (XS) Control de calidad para el analizador **ProCyte Dx***
- e-CHECK*** (XS) Qualitätskontrolle für das **ProCyte Dx*** Analysegerät
- e-CHECK*** (XS) Controllo qualità per l'analizzatore **ProCyte Dx***
- e-CHECK*** (XS) IDEXXプロサイトDx*の精度管理

L1 (Low / Bas / Bajo / Niedrig / Basso / 低濃度)

LOT 52720804  12/21/25

Parameter Paramètre Parámetro Parameter Parametro 項目	Target Valeur cible Objetivo Ergebnis Valore ideale 目標値	Expected Range Intervalle attendu Rango esperado Referenzbereich Intervallo atteso 基準範囲
RBC/GR (10 ⁶ /μL)	2.25	2.10 - 2.40
HCT (%)	16.6	13.6 - 19.6
HGB (g/dL)	5.5	5.2 - 5.8
MCV/VGM (fL)	73.8	66.3 - 81.3
MCH/TCMH (pg)	24.4	20.6 - 28.2
MCHC/CCMH (g/dL)	33.1	21.9 - 44.3
PLT (K/μL)	53	23 - 83
RDW-SD (fL)	45.5	36.5 - 54.5
RDW-CV (%)	17.7	14.2 - 21.2
PDW (fL)	9.2	5.2 - 13.2
PCT (%)	0.05	0.03 - 0.07
MPV (fL)	8.5	5.5 - 11.5
P-LCR (%)	10.7	3.2 - 18.2
WBC/GB (K/μL)	3.04	2.64 - 3.44
NEUT# (K/μL)	1.37	0.96 - 1.78
LYMPH# (K/μL)	1.09	0.69 - 1.49
MONO# (K ⁹ /μL)	0.29	0.00 - 0.59
EO# (K/μL)	0.29	0.09 - 0.49
RBC-O (10 ⁶ /μL)	2.25	1.85 - 2.65
PLT-O (K/μL)	43	13 - 73
RETIC (K/μL)	149.1	109.1 - 189.1
% RETIC (%)	6.63	4.93 - 8.33

L2 (Normal / Normale / 正常濃度)

LOT 52720805  12/21/25

Parameter Paramètre Parámetro Parameter Parametro 項目	Target Valeur cible Objetivo Ergebnis Valore ideale 目標値	Expected Range Intervalle attendu Rango esperado Referenzbereich Intervallo atteso 基準範囲
RBC/GR (10 ⁶ /μL)	4.37	4.10 - 4.64
HCT (%)	36.6	33.1 - 40.1
HGB (g/dL)	12.2	11.7 - 12.7
MCV/VGM (fL)	83.8	75.8 - 91.8
MCH/TCMH (pg)	27.9	24.2 - 31.6
MCHC/CCMH (g/dL)	33.3	27.2 - 39.4
PLT (K/μL)	199	154 - 244
RDW-SD (fL)	48.2	41.2 - 55.2
RDW-CV (%)	16.9	13.9 - 19.9
PDW (fL)	9.4	6.4 - 12.4
PCT (%)	0.19	0.11 - 0.27
MPV (fL)	9.4	6.4 - 12.4
P-LCR (%)	14.0	8.1 - 19.9
WBC/GB (K/μL)	7.18	6.43 - 7.93
NEUT# (K/μL)	3.47	2.41 - 4.53
LYMPH# (K/μL)	2.28	1.66 - 2.90
MONO# (K ⁹ /μL)	0.69	0.00 - 1.45
EO# (K/μL)	0.74	0.38 - 1.10
RBC-O (10 ⁶ /μL)	4.20	3.50 - 4.90
PLT-O (K/μL)	183	133 - 233
RETIC (K/μL)	120.5	90.5 - 150.5
% RETIC (%)	2.76	1.96 - 3.56