

- e-CHECK* (XS)** Quality Control for the **ProCyte Dx*** Analyzer
- e-CHECK* (XS)** contrôle qualité pour l'analyseur **ProCyte Dx***
- e-CHECK* (XS)** Control de calidad para el analizador **ProCyte Dx***
- e-CHECK* (XS)** Qualitätskontrolle für das **ProCyte Dx*** Analysegerät
- e-CHECK* (XS)** Controllo qualita' per l'analizzatore **ProCyte Dx***
- e-CHECK* (XS)** IDEXXプロサイトDx*の精度管理

L1 (Low / Bas / Bajo / Niedrig / Basso / 低濃度)

LOT 61310804



08/02/26

L2 (Normal / Normale / 正常濃度)

LOT 61310805



08/02/26

Parameter Paramètre Parámetro Parameter Parametro 項目	Target Valeur cible Objetivo Ergebnis Valore ideale 目標値	Expected Range Intervalle attendu Rango esperado Referenzbereich Intervallo atteso 基準範囲
RBC/GR (10 ⁹ /μL)	2.32	2.17 - 2.47
HCT (%)	16.5	13.5 - 19.5
HGB (g/dL)	5.5	5.2 - 5.8
MCV/VGM (fL)	71.2	63.7 - 78.7
MCH/TCMH (pg)	23.7	19.9 - 27.5
MCHC/CCMH (g/dL)	33.3	22.1 - 44.5
PLT (K/μL)	56	26 - 86
RDW-SD (fL)	47.3	38.3 - 56.3
RDW-CV (%)	19.1	15.6 - 22.6
PDW (fL)	8.8	4.8 - 12.8
PCT (%)	0.05	0.03 - 0.07
MPV (fL)	8.5	5.5 - 11.5
P-LCR (%)	10.3	2.8 - 17.8
WBC/GB (K/μL)	3.02	2.62 - 3.42
NEUT# (K/μL)	1.39	0.98 - 1.80
LYMPH# (K/μL)	1.06	0.66 - 1.46
MONO# (K ⁹ /μL)	0.28	0.00 - 0.58
EO# (K/μL)	0.29	0.09 - 0.49
RBC-O (10 ⁹ /μL)	2.34	1.94 - 2.74
PLT-O (K/μL)	43	13 - 73
RETIC (K/μL)	148.9	108.9 - 188.9
% RETIC (%)	6.42	4.72 - 8.12

Parameter Paramètre Parámetro Parameter Parametro 項目	Target Valeur cible Objetivo Ergebnis Valore ideale 目標値	Expected Range Intervalle attendu Rango esperado Referenzbereich Intervallo atteso 基準範囲
RBC/GR (10 ⁹ /μL)	4.34	4.07 - 4.61
HCT (%)	35.1	31.6 - 38.6
HGB (g/dL)	11.7	11.2 - 12.2
MCV/VGM (fL)	80.8	72.8 - 88.8
MCH/TCMH (pg)	27.0	23.3 - 30.7
MCHC/CCMH (g/dL)	33.3	27.2 - 39.4
PLT (K/μL)	212	167 - 257
RDW-SD (fL)	46.4	39.4 - 53.4
RDW-CV (%)	17.1	14.1 - 20.1
PDW (fL)	9.3	6.3 - 12.3
PCT (%)	0.20	0.12 - 0.28
MPV (fL)	9.2	6.2 - 12.2
P-LCR (%)	12.8	6.9 - 18.7
WBC/GB (K/μL)	7.62	6.87 - 8.37
NEUT# (K/μL)	3.80	2.74 - 4.86
LYMPH# (K/μL)	2.24	1.62 - 2.86
MONO# (K ⁹ /μL)	0.79	0.03 - 1.55
EO# (K/μL)	0.79	0.43 - 1.15
RBC-O (10 ⁹ /μL)	4.24	3.54 - 4.94
PLT-O (K/μL)	169	119 - 219
RETIC (K/μL)	120.4	90.4 - 150.4
% RETIC (%)	2.77	1.97 - 3.57