


- e-CHECK* (XS)** Quality Control for the **ProCyte Dx*** Analyzer
- e-CHECK* (XS)** contrôle qualité pour l'analyseur **ProCyte Dx***
- e-CHECK* (XS)** Control de calidad para el analizador **ProCyte Dx***
- e-CHECK* (XS)** Qualitätskontrolle für das **ProCyte Dx*** Analysegerät
- e-CHECK* (XS)** Controllo qualita' per l'analizzatore **ProCyte Dx***
- e-CHECK* (XS)** IDEXXプロサイトDx*の精度管理


L1 (Low / Bas / Bajo / Niedrig / Basso / 低濃度)

LOT 61870804

 09/27/26

L2 (Normal / Normale / 正常濃度)

LOT 61870805

 09/27/26

Parameter Paramètre Parámetro Parameter Parametro 項目	Target Valeur cible Objetivo Ergebnis Valore ideale 目標値	Expected Range Intervalle attendu Rango esperado Referenzbereich Intervallo atteso 基準範囲
RBC/GR (10 ⁹ /μL)	2.22	2.07 - 2.37
HCT (%)	15.6	12.6 - 18.6
HGB (g/dL)	5.1	4.8 - 5.4
MCV/VGM (fL)	70.4	62.9 - 77.9
MCH/TCMH (pg)	23.0	19.2 - 26.8
MCHC/CCMH (g/dL)	32.7	21.5 - 43.9
PLT (K/μL)	54	24 - 84
RDW-SD (fL)	44.2	35.2 - 53.2
RDW-CV (%)	18.0	14.5 - 21.5
PDW (fL)	8.9	4.9 - 12.9
PCT (%)	0.04	0.02 - 0.06
MPV (fL)	8.3	5.3 - 11.3
P-LCR (%)	8.9	1.4 - 16.4
WBC/GB (K/μL)	2.96	2.56 - 3.36
NEUT# (K/μL)	1.34	0.93 - 1.75
LYMPH# (K/μL)	1.06	0.66 - 1.46
MONO# (K ⁹ /μL)	0.27	0.00 - 0.57
EO# (K/μL)	0.29	0.09 - 0.49
RBC-O (10 ⁹ /μL)	2.26	1.86 - 2.66
PLT-O (K/μL)	42	12 - 72
RETIC (K/μL)	137.5	97.5 - 177.5
% RETIC (%)	6.19	4.49 - 7.89

Parameter Paramètre Parámetro Parameter Parametro 項目	Target Valeur cible Objetivo Ergebnis Valore ideale 目標値	Expected Range Intervalle attendu Rango esperado Referenzbereich Intervallo atteso 基準範囲
RBC/GR (10 ⁹ /μL)	4.30	4.03 - 4.57
HCT (%)	35.8	32.3 - 39.3
HGB (g/dL)	12.0	11.5 - 12.5
MCV/VGM (fL)	83.3	75.3 - 91.3
MCH/TCMH (pg)	27.9	24.2 - 31.6
MCHC/CCMH (g/dL)	33.5	27.4 - 39.6
PLT (K/μL)	215	170 - 260
RDW-SD (fL)	45.8	38.8 - 52.8
RDW-CV (%)	16.1	13.1 - 19.1
PDW (fL)	9.1	6.1 - 12.1
PCT (%)	0.19	0.11 - 0.27
MPV (fL)	9.0	6.0 - 12.0
P-LCR (%)	11.1	5.2 - 17.0
WBC/GB (K/μL)	7.04	6.29 - 7.79
NEUT# (K/μL)	3.31	2.25 - 4.37
LYMPH# (K/μL)	2.16	1.54 - 2.78
MONO# (K ⁹ /μL)	0.88	0.12 - 1.64
EO# (K/μL)	0.69	0.33 - 1.05
RBC-O (10 ⁹ /μL)	4.24	3.54 - 4.94
PLT-O (K/μL)	172	122 - 222
RETIC (K/μL)	111.3	81.3 - 141.3
% RETIC (%)	2.59	1.79 - 3.39