

- e-CHECK*** (XS) Quality Control for the **ProCyte Dx*** Analyzer
- e-CHECK*** (XS) contrôle qualité pour l'analyseur **ProCyte Dx***
- e-CHECK*** (XS) Control de calidad para el analizador **ProCyte Dx***
- e-CHECK*** (XS) Qualitätskontrolle für das **ProCyte Dx*** Analysegerät
- e-CHECK*** (XS) Controllo qualità per l'analizzatore **ProCyte Dx***
- e-CHECK*** (XS) IDEXXプロサイトDx*の精度管理

L1 (Low / Bas / Bajo / Niedrig / Basso / 低濃度)

LOT 62430804  11/22/26

Parameter Paramètre Parámetro Parameter Parametro 項目	Target Valeur cible Objetivo Ergebnis Valore ideale 目標値	Expected Range Intervalle attendu Rango esperado Referenzbereich Intervallo atteso 基準範囲
RBC/GR (10 ⁶ /μL)	2.23	2.08 - 2.38
HCT (%)	16.7	13.7 - 19.7
HGB (g/dL)	5.6	5.3 - 5.9
MCV/VGM (fL)	75.0	67.5 - 82.5
MCH/TCMH (pg)	25.1	21.3 - 28.9
MCHC/CCMH (g/dL)	33.5	22.3 - 44.7
PLT (K/μL)	55	25 - 85
RDW-SD (fL)	47.3	38.3 - 56.3
RDW-CV (%)	18.0	14.5 - 21.5
PDW (fL)	9.1	5.1 - 13.1
PCT (%)	0.05	0.03 - 0.07
MPV (fL)	8.3	5.3 - 11.3
P-LCR (%)	10.0	2.5 - 17.5
WBC/GB (K/μL)	3.04	2.64 - 3.44
NEUT# (K/μL)	1.34	0.93 - 1.75
LYMPH# (K/μL)	1.11	0.71 - 1.51
MONO# (K ⁹ /μL)	0.30	0.00 - 0.60
EO# (K/μL)	0.29	0.09 - 0.49
RBC-O (10 ⁶ /μL)	2.27	1.87 - 2.67
PLT-O (K/μL)	40	10 - 70
RETIC (K/μL)	145.7	105.7 - 185.7
% RETIC (%)	6.53	4.83 - 8.23

L2 (Normal / Normale / 正常濃度)

LOT 62430805  11/22/26

Parameter Paramètre Parámetro Parameter Parametro 項目	Target Valeur cible Objetivo Ergebnis Valore ideale 目標値	Expected Range Intervalle attendu Rango esperado Referenzbereich Intervallo atteso 基準範囲
RBC/GR (10 ⁶ /μL)	4.31	4.04 - 4.58
HCT (%)	35.8	32.3 - 39.3
HGB (g/dL)	12.1	11.6 - 12.6
MCV/VGM (fL)	83.0	75.0 - 91.0
MCH/TCMH (pg)	28.1	24.4 - 31.8
MCHC/CCMH (g/dL)	33.8	27.7 - 39.9
PLT (K/μL)	213	168 - 258
RDW-SD (fL)	46.7	39.7 - 53.7
RDW-CV (%)	16.5	13.5 - 19.5
PDW (fL)	9.4	6.4 - 12.4
PCT (%)	0.19	0.11 - 0.27
MPV (fL)	9.0	6.0 - 12.0
P-LCR (%)	12.2	6.3 - 18.1
WBC/GB (K/μL)	7.10	6.35 - 7.85
NEUT# (K/μL)	3.42	2.36 - 4.48
LYMPH# (K/μL)	2.23	1.61 - 2.85
MONO# (K ⁹ /μL)	0.73	0.00 - 1.49
EO# (K/μL)	0.72	0.36 - 1.08
RBC-O (10 ⁶ /μL)	4.23	3.53 - 4.93
PLT-O (K/μL)	170	120 - 220
RETIC (K/μL)	115.1	85.1 - 145.1
% RETIC (%)	2.67	1.87 - 3.47